



TENNESSEE DEPARTMENT OF HEALTH
 DIVISION OF HEALTH LICENSURE AND REGULATION
OFFICE OF EMERGENCY MEDICAL SERVICES
 665 MAINSTREAM DRIVE, 2ND FLOOR
 NASHVILLE, TN 37243
 TELEPHONE: 615-741-2584

**NEONATAL TRANSPORT
 VEHICLE INSPECTION SUPPLEMENT**

Date: _____

Service: _____ Permit No.: _____

Pass **Fail**

Inspector's Signature: _____

EQUIPMENT

A U (A = Acceptable U = Unacceptable)

- Transport Incubator
- Thermometer
- Blanket (insulating blanket or chemically activated warm pack approved for neonate use)

Above equipment maintained in neutral thermal environment.

OXYGEN DELIVERY & MONITORING

A U (A = Acceptable U = Unacceptable)

- Oxygen Tanks
- Air Tanks
- Pressure Gauges
- Flow-Meters
- Oxygen Analyzer
- Oxygen Blender
- Oxygen Tubing and Adapters
- Oxygen Hood or Nasal Cannula
- Neonatal Oxygen Mask
- Neonatal Resuscitation BVM (with Manometer)
- Continuous Positive Airway Pressure Apparatus
- Mechanical Ventilator
- Non-Invasive Oxygen Monitor (i.e. Pulse Ox)

AIRWAY / GASTRIC DECOMPRESSION

A U (A = Acceptable U = Unacceptable)

- Bulb Syringe
- Regulated Suction w/Gauge
- Suction Catheters (#6, 8, 10 Fr)
- Feeding Tube (# 8 Fr.) w/20 ml Syringe
- Replogle Tube (#8 Fr.)

MONITORING (VITAL SIGNS)

A U (A = Acceptable U = Unacceptable)

- Continuous Heart-Rate Monitor
- Neonatal Stethoscope
- Body Temperature Monitor
- Non-Invasive & Invasive BP Monitor
- Glucometer or Equivalent

PUMPS

A U (A = Acceptable U = Unacceptable)

- IV Pump
(portable, battery powered, failsafe & calibrated)

ENDOTRACHEAL INTUBATION

A U (A = Acceptable U = Unacceptable)

- Laryngoscope Handle with Blades (00, 0, 1)
- Spare Laryngoscope Bulbs
- Spare Batteries
- ET Tubes (# 2.5, 3.0, 3.5, 4.0 mm I.D.)
- Disposable Stylette (# 6 Fr.)
- Adhesive Tape
- Scissors

